

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00834	Calib. Date: 09/14/2018
Certification File No.:	00835	Calib. No.: 00028
Linearity File No.:	00836	Cert. Date: 09/14/2018
Solution File No.:	00845	Cert. No.: 00025
Sequential File No.:	00845	Lin. Date: 09/14/2018
		Lin. No.: 00026
		Soln. Date: 10/20/2018
		Soln. No.: 00180
		File Date: 10/20/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0039
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0341

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:44D		
Control 1 EC	0.103%	10:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.104%	10:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45D		
Control 2 EC	0.103%	10:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47D		
Control 3 EC	0.103%	10:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: *BSK* *P41 Brad Kerney 131*

Badge No.: 131

Date: 10/20/2018