

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00834	Calib. Date: 09/14/2018
Certification File No.:	00835	Calib. No.: 00028
Linearity File No.:	00836	Cert. Date: 09/14/2018
Solution File No.:	00854	Cert. No.: 00025
Sequential File No.:	00854	Lin. Date: 09/14/2018
		Lin. No.: 00026
		Soln. Date: 11/22/2018
		Soln. No.: 00181
		File Date: 11/22/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0039
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0168

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:44S		
Control 1 EC	0.098%	01:45S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:46S		
Control 2 EC	0.098%	01:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:47S		
Control 3 EC	0.099%	01:48S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: _____

Badge No.: 130

Date: 11/22/2018