

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: DEAL POLICE
 Calibration File No.: 00834
 Certification File No.: 00835
 Linearity File No.: 00836
 Solution File No.: 00858
 Sequential File No.: 00858

Serial No.: ARWM-0030

Calib. Date: 09/14/2018 Calib. No.: 00028
 Cert. Date: 09/14/2018 Cert. No.: 00025
 Lin. Date: 09/14/2018 Lin. No.: 00026
 Soln. Date: 11/30/2018 Soln. No.: 00182
 File Date: 11/30/2018

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 17420

Model No.: CU-34

Serial No.: DDXA S3-0039
 Expires: 12/11/2019
 Bottle No.: 0406

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:41S		
Control 1 EC	0.000%	21:42S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.000%	21:42S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	21:42S		
Control 2 EC	-.---%	21:43S	33.9°C	CTRL GAS SUPPLY
Control 2 IR	-.---%	21:43S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: _____

Badge No.: 130

Date: 11/30/2018