

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00788	Calib. Date: 04/12/2018
Certification File No.:	00789	Calib. No.: 00027
Linearity File No.:	00790	Cert. Date: 04/12/2018
Solution File No.:	00830	Cert. No.: 00024
Sequential File No.:	00830	Lin. Date: 04/12/2018
		Lin. No.: 00025
		Soln. Date: 08/26/2018
		Soln. No.: 00177
		File Date: 08/26/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0038
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0712

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:04D		
Control 1 EC	0.098%	18:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	18:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:05D		
Control 2 EC	0.099%	18:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:07D		
Control 3 EC	0.099%	18:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: \_\_\_\_\_

Badge No.: 130

Date: 08/26/2018