

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0030
Location: DEAL POLICE
Calibration File No.: 00834 Calib. Date: 09/14/2018 Calib. No.: 00028
Certification File No.: 00835 Cert. Date: 09/14/2018 Cert. No.: 00025
Linearity File No.: 00836 Lin. Date: 09/14/2018 Lin. No.: 00026
Solution File No.: 00882 Soln. Date: 01/12/2019 Soln. No.: 00187
Sequential File No.: 00882 File Date: 01/12/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0039
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 1210

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:39S		
Control 1 EC	0.096%	19:39S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:40S		
Control 2 EC	0.097%	19:41S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:42S		
Control 3 EC	0.097%	19:43S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: 

Badge No.: 130

Date: 01/12/2019