

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00925	Calib. Date: 07/17/2019
Certification File No.:	00926	Calib. No.: 00030
Linearity File No.:	00927	Cert. Date: 07/17/2019
Solution File No.:	00936	Cert. No.: 00027
Sequential File No.:	00936	Lin. Date: 07/17/2019
		Lin. No.: 00028
		Soln. Date: 10/13/2019
		Soln. No.: 00200
		File Date: 10/13/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0039
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 0576

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:07D		
Control 1 EC	0.100%	00:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08D		
Control 2 EC	0.100%	00:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:10D		
Control 3 EC	0.099%	00:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature:  131

Badge No.: 131

Date: 10/13/2019