

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0030
Location: DEAL POLICE
Calibration File No.: 00925 Calib. Date: 07/17/2019 Calib. No.: 00030
Certification File No.: 00926 Cert. Date: 07/17/2019 Cert. No.: 00027
Linearity File No.: 00927 Lin. Date: 07/17/2019 Lin. No.: 00028
Solution File No.: 00941 Soln. Date: 12/04/2019 Soln. No.: 00202
Sequential File No.: 00941 File Date: 12/04/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0039
Control Solution %: 0.100% Expires: 02/11/2021
Solution Control Lot: 19060 Bottle No.: 0403

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:19S		
Control 1 EC	0.098%	01:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:21S		
Control 2 EC	0.097%	01:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22S		
Control 3 EC	0.099%	01:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Brad Kerney

131 DPD

Badge No.: 131

Date: 12/04/2019