

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: DEAL POLICE
 Calibration File No.: 00925
 Certification File No.: 00926
 Linearity File No.: 00927
 Solution File No.: 00945
 Sequential File No.: 00945

Serial No.: ARWM-0030

Calib. Date: 07/17/2019 Calib. No.: 00030
 Cert. Date: 07/17/2019 Cert. No.: 00027
 Lin. Date: 07/17/2019 Lin. No.: 00028
 Soln. Date: 12/11/2019 Soln. No.: 00205
 File Date: 12/11/2019

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 19060

Model No.: CU-34

Serial No.: DDXA S3-0039
 Expires: 02/11/2021
 Bottle No.: 0517

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:12S		*** TEST PASSED ***
Control 1 EC	0.098%	22:12S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:13S		*** TEST PASSED ***
Control 2 EC	0.098%	22:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:15S		*** TEST PASSED ***
Control 3 EC	0.098%	22:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. BT 131 DPD

Badge No.: 131

Date: 12/11/2019