

# Calibrating Unit

## New Standard Solution Report

|                         |                       |                          |
|-------------------------|-----------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARWM-0030    |
| Location:               | DEAL POLICE           |                          |
| Calibration File No.:   | 00925                 | Calib. Date: 07/17/2019  |
| Certification File No.: | 00926                 | Calib. No.: 00030        |
| Linearity File No.:     | 00927                 | Cert. Date: 07/17/2019   |
| Solution File No.:      | 00948                 | Cert. No.: 00027         |
| Sequential File No.:    | 00948                 | Lin. Date: 07/17/2019    |
|                         |                       | Lin. No.: 00028          |
|                         |                       | Soln. Date: 12/17/2019   |
|                         |                       | Soln. No.: 00206         |
|                         |                       | File Date: 12/17/2019    |
| <br>                    |                       |                          |
| Calibrating Unit:       | WET                   | Model No.: CU-34         |
| Control Solution %:     | 0.100%                | Serial No.: DDXA S3-0039 |
| Solution Control Lot:   | 19060                 | Expires: 02/11/2021      |
|                         |                       | Bottle No.: 0575         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 22:22S |                |                     |
| Control 1 EC      | 0.096% | 22:23S | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.097% | 22:23S | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:23S |                |                     |
| Control 2 EC      | 0.097% | 22:24S | 33.9°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.097% | 22:24S | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:25S |                |                     |
| Control 3 EC      | 0.097% | 22:26S | 33.9°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.098% | 22:26S | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:26S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Bk

*131 DPD*

Badge No.: 131

Date: 12/17/2019