

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00834	Calib. Date: 09/14/2018
Certification File No.:	00835	Calib. No.: 00028
Linearity File No.:	00836	Cert. Date: 09/14/2018
Solution File No.:	00884	Cert. No.: 00025
Sequential File No.:	00884	Lin. Date: 09/14/2018
		Lin. No.: 00026
		Soln. Date: 02/01/2019
		Soln. No.: 00188
		File Date: 02/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0039
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0692

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:32S		
Control 1 EC	0.099%	17:33S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:33S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:33S		
Control 2 EC	0.099%	17:34S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:35S		
Control 3 EC	0.099%	17:36S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: \_\_\_\_\_

*Pt. B Kerney 131 DPD*

Badge No.: 131

Date: 02/01/2019