Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWM-0030
Location:	DEAL POLICE				
Calibration File No.:	00834	Calib. Date:	09/14/2018	Calib. No.:	00028
Certification File No.:	00835	Cert. Date:	09/14/2018	Cert. No.:	00025
Linearity File No.:	00836	Lin. Date:	09/14/2018	Lin. No.:	00026
Solution File No.:	00887	Soln. Date:	02/09/2019	Soln. No.:	00189
Sequential File No.:	00887	File Date:	02/09/2019		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXA S3-0039
Control Solution %:	0.100%			Expires:	09/17/2020
Solution Control Lot:	18290			Bottle No.:	0854

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:31S		
Control 1 EC	0.075%	02:32S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.077%	02:32S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	02:33S		
Control 2 EC	-.---%	02:33S	33.9°C	CTRL GAS SUPPLY
Control 2 IR	-.---%	02:33S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:---		
Control 3 EC	-.---%	--:---	N/A	
Control 3 IR	-.---%	--:---	N/A	
Ambient Air Blank	-.---%	--:---		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO First Name: JONATHAN MI: L

Signature: #130 Badge No.: 130

Date: 02/09/2019