

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00891	Calib. Date: 02/28/2019
Certification File No.:	00892	Calib. No.: 00029
Linearity File No.:	00893	Cert. Date: 02/28/2019
Solution File No.:	00908	Cert. No.: 00026
Sequential File No.:	00908	Lin. Date: 02/28/2019
		Lin. No.: 00027
		Soln. Date: 05/02/2019
		Soln. No.: 00193
		File Date: 05/02/2019

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0038
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0882

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:26D		
Control 1 EC	-.---%	07:27D	20.7°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	07:27D	20.7°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KERNEY

First Name: BRAD

MI: S

Signature:  151 DPD

Badge No.: 131

Date: 05/02/2019