

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0030
Location: DEAL POLICE
Calibration File No.: 00891 Calib. Date: 02/28/2019 Calib. No.: 00029
Certification File No.: 00892 Cert. Date: 02/28/2019 Cert. No.: 00026
Linearity File No.: 00893 Lin. Date: 02/28/2019 Lin. No.: 00027
Solution File No.: 00909 Soln. Date: 05/02/2019 Soln. No.: 00194
Sequential File No.: 00909 File Date: 05/02/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0038
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0882

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:53D		
Control 1 EC	0.101%	20:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:54D		
Control 2 EC	0.100%	20:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55D		
Control 3 EC	0.100%	20:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: 

131 OPD

Badge No.: 131

Date: 05/02/2019