

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00891	Calib. Date: 02/28/2019
Certification File No.:	00892	Calib. No.: 00029
Linearity File No.:	00893	Cert. Date: 02/28/2019
Solution File No.:	00918	Cert. No.: 00026
Sequential File No.:	00918	Lin. Date: 02/28/2019
		Lin. No.: 00027
		Soln. Date: 06/02/2019
		Soln. No.: 00195
		File Date: 06/02/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0038
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 1091

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:56D		
Control 1 EC	0.100%	03:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:57D		
Control 2 EC	0.100%	03:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	03:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:59D		
Control 3 EC	0.101%	04:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: _____

Badge No.: 130

Date: 06/02/2019