

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWM-0030  
Location: DEAL POLICE  
Calibration File No.: 00891 Calib. Date: 02/28/2019 Calib. No.: 00029  
Certification File No.: 00892 Cert. Date: 02/28/2019 Cert. No.: 00026  
Linearity File No.: 00893 Lin. Date: 02/28/2019 Lin. No.: 00027  
Solution File No.: 00923 Soln. Date: 07/06/2019 Soln. No.: 00196  
Sequential File No.: 00923 File Date: 07/06/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0038  
Control Solution %: 0.100% Expires: 09/17/2020  
Solution Control Lot: 18290 Bottle No.: 0981

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:29D		
Control 1 EC	0.100%	23:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:31D		
Control 2 EC	0.100%	23:31D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:32D		
Control 3 EC	0.100%	23:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:33D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature:  131 DPD

Badge No.: 131

Date: 07/06/2019