

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0030
Location: DEAL POLICE
Calibration File No.: 00925 Calib. Date: 07/17/2019 Calib. No.: 00030
Certification File No.: 00926 Cert. Date: 07/17/2019 Cert. No.: 00027
Linearity File No.: 00927 Lin. Date: 07/17/2019 Lin. No.: 00028
Solution File No.: 00932 Soln. Date: 08/25/2019 Soln. No.: 00198
Sequential File No.: 00932 File Date: 08/25/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0039
Control Solution %: 0.100% Expires: 02/11/2021
Solution Control Lot: 19060 Bottle No.: 0574

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:16D		
Control 1 EC	0.098%	17:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17D		
Control 2 EC	0.098%	17:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		
Control 3 EC	0.098%	17:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature:  #130

Badge No.: 130

Date: 08/25/2019