

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: DEAL POLICE
 Calibration File No.: 00950
 Certification File No.: 00951
 Linearity File No.: 00952
 Solution File No.: 00963
 Sequential File No.: 00963

Serial No.: ARWM-0030

Calib. Date: 01/10/2020 Calib. No.: 00031
 Cert. Date: 01/10/2020 Cert. No.: 00028
 Lin. Date: 01/10/2020 Lin. No.: 00029
 Soln. Date: 02/25/2020 Soln. No.: 00210
 File Date: 02/25/2020

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 19300

Model No.: CU-34

Serial No.: DDXA S3-0038
 Expires: 10/30/2021
 Bottle No.: 1166

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:57S		
Control 1 EC	0.097%	17:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:59S		
Control 2 EC	0.098%	18:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	18:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:00S		
Control 3 EC	0.098%	18:01S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SHARIN

First Name: MATTHEW

MI: J

Signature: _____



Badge No.: 123

Date: 02/25/2020