

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00950	Calib. Date: 01/10/2020
Certification File No.:	00951	Calib. No.: 00031
Linearity File No.:	00952	Cert. Date: 01/10/2020
Solution File No.:	00956	Cert. No.: 00028
Sequential File No.:	00956	Lin. Date: 01/10/2020
		Lin. No.: 00029
		Soln. Date: 01/18/2020
		Soln. No.: 00208
		File Date: 01/18/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0038
Solution Control Lot:	19060	Expires: 02/11/2021
		Bottle No.: 0404

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:44S		
Control 1 EC	0.097%	16:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:45S		
Control 2 EC	0.097%	16:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47S		
Control 3 EC	0.098%	16:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. BSA 131 DPD

Badge No.: 131

Date: 01/18/2020