

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	00950	Calib. Date:	01/10/2020
Certification File No.:	00951	Cert. No.:	00031
Linearity File No.:	00952	Cert. Date:	01/10/2020
Solution File No.:	00961	Lin. Date:	01/10/2020
Sequential File No.:	00961	Soln. Date:	02/15/2020
		File Date:	02/15/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0038
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0579

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:27S		
Control 1 EC	0.096%	06:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:28S		
Control 2 EC	0.096%	06:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	06:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:30S		
Control 3 EC	0.096%	06:31S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	06:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KERNEY	131	First Name: BRAD	MI: S
Signature: <u>Sgt. Kerney</u>		Badge No.: 131	Date: 02/15/2020