

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE	Calib. Date:	06/29/2020
Calibration File No.:	00969	Calib. No.:	00032
Certification File No.:	00970	Cert. Date:	06/29/2020
Linearity File No.:	00971	Lin. Date:	06/29/2020
Solution File No.:	00990	Soln. Date:	10/21/2020
Sequential File No.:	00990	File Date:	10/21/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	4
Solution Control Lot:	20390	Expires:	09/02/2022
		Bottle No.:	1450

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:06D		
Control 1 EC	0.094%	14:06D	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.094%	14:06D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:07D		
Control 2 EC	0.094%	14:08D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.095%	14:08D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:09D		
Control 3 EC	0.094%	14:10D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.095%	14:10D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:11D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: _____

Badge No.: 130

Date: 10/21/2020