

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030		
Location:	DEAL POLICE				
Calibration File No.:	00969	Calib. Date:	06/29/2020	Calib. No.:	00032
Certification File No.:	00970	Cert. Date:	06/29/2020	Cert. No.:	00029
Linearity File No.:	00971	Lin. Date:	06/29/2020	Lin. No.:	00030
Solution File No.:	00994	Soln. Date:	11/16/2020	Soln. No.:	00228
Sequential File No.:	00994	File Date:	11/16/2020		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	4
Control Solution %:	0.100%			Expires:	09/02/2022
Solution Control Lot:	20390			Bottle No.:	1438

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:03S		
Control 1 EC	-.---%	01:04S	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	01:04S	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: 

Badge No.: 130

Date: 11/16/2020