

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	00950	Calib. Date:	01/10/2020
Certification File No.:	00951	Cert. No.:	00031
Linearity File No.:	00952	Lin. Date:	01/10/2020
Solution File No.:	00966	Cert. No.:	00028
Sequential File No.:	00966	Lin. No.:	00029
		Soln. Date:	05/22/2020
		Soln. No.:	00211
		File Date:	05/22/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0038
Solution Control Lot:	19300	Expires:	10/30/2021
		Bottle No.:	1110

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:20D		
Control 1 EC	0.101%	19:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:22D		
Control 2 EC	0.101%	19:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:23D		
Control 3 EC	0.101%	19:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Brad Kerney 131 DPD

Badge No.: KERNEY

Date: 05/22/2020