

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	00950	Calib. Date:	01/10/2020
Certification File No.:	00951	Cert. Date:	01/10/2020
Linearity File No.:	00952	Lin. Date:	01/10/2020
Solution File No.:	00968	Soln. Date:	06/25/2020
Sequential File No.:	00968	File Date:	06/25/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0038
Solution Control Lot:	19300	Expires:	10/30/2021
		Bottle No.:	1178

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:56D		
Control 1 EC	0.101%	19:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:58D		
Control 2 EC	0.100%	19:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:59D		
Control 3 EC	0.101%	20:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DOWLING

First Name: NICHOLAS

MI: C

Signature: _____

Badge No.: 136

Date: 06/25/2020