

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00969	Calib. Date: 06/29/2020
Certification File No.:	00970	Calib. No.: 00032
Linearity File No.:	00971	Cert. Date: 06/29/2020
Solution File No.:	00977	Cert. No.: 00029
Sequential File No.:	00977	Lin. Date: 06/29/2020
		Lin. No.: 00030
		Soln. Date: 08/24/2020
		Soln. No.: 00217
		File Date: 08/24/2020

Calibrating Unit: : WET	Model No.: CU-34	Serial No.: 4
Control Solution %: 0.100%		Expires: 10/30/2021
Solution Control Lot: 19300		Bottle No.: 1179

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:10D		
Control 1 EC	0.099%	00:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:11D		
Control 2 EC	0.100%	00:12D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:13D		
Control 3 EC	0.100%	00:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Brad Kerney 131

Badge No.: 131

Date: 08/24/2020