

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	00969	Calib. Date:	06/29/2020
Certification File No.:	00970	Cert. No.:	00032
Linearity File No.:	00971	Cert. Date:	06/29/2020
Solution File No.:	00974	Lin. Date:	06/29/2020
Sequential File No.:	00974	Lin. No.:	00030
		Soln. Date:	08/03/2020
		Soln. No.:	00215
		File Date:	08/03/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0039
Solution Control Lot:	19300	Expires:	10/30/2021
		Bottle No.:	1118

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:24D		
Control 1 EC	-.---%	01:25D	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	01:25D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. B Kerney 131 DPD

Badge No.: 131

Date: 08/03/2020