

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	00969	Calib. Date:	06/29/2020
Certification File No.:	00970	Cert. No.:	00032
Linearity File No.:	00971	Cert. Date:	06/29/2020
Solution File No.:	00975	Lin. Date:	06/29/2020
Sequential File No.:	00975	Lin. No.:	00030
		Soln. Date:	08/03/2020
		Soln. No.:	00216
		File Date:	08/03/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	4
Solution Control Lot:	19300	Expires:	10/30/2021
		Bottle No.:	1118

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:44D		
Control 1 EC	0.098%	06:44D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	06:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:45D		
Control 2 EC	0.098%	06:46D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:47D		
Control 3 EC	0.098%	06:47D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	06:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: *Sgt B Kerney* 131

Badge No.: 131

Date: 08/03/2020