

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00969	Calib. Date: 06/29/2020
Certification File No.:	00970	Calib. No.: 00032
Linearity File No.:	00971	Cert. Date: 06/29/2020
Solution File No.:	00983	Cert. No.: 00029
Sequential File No.:	00983	Lin. Date: 06/29/2020
		Lin. No.: 00030
		Soln. Date: 09/24/2020
		Soln. No.: 00220
		File Date: 09/24/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: 4
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 1122

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:26D		
Control 1 EC	0.097%	16:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	16:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28D		
Control 2 EC	0.098%	16:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29D		
Control 3 EC	0.099%	16:30D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: _____

Badge No.: 130

Date: 09/24/2020