

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWM-0030  
Location: DEAL POLICE  
Calibration File No.: 00995 Calib. Date: 11/30/2020 Calib. No.: 00033  
Certification File No.: 00996 Cert. Date: 11/30/2020 Cert. No.: 00030  
Linearity File No.: 00997 Lin. Date: 11/30/2020 Lin. No.: 00031  
Solution File No.: 01005 Soln. Date: 02/03/2021 Soln. No.: 00232  
Sequential File No.: 01005 File Date: 02/03/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0038  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 0234

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:56S		
Control 1 EC	0.096%	09:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		
Control 2 EC	0.096%	09:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	09:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00S		
Control 3 EC	0.097%	10:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: DOWLING

First Name: NICHOLAS

MI: C

Signature: 

Badge No.: 136

Date: 02/03/2021