

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0030
Location: DEAL POLICE
Calibration File No.: 00995 Calib. Date: 11/30/2020 Calib. No.: 00033
Certification File No.: 00996 Cert. Date: 11/30/2020 Cert. No.: 00030
Linearity File No.: 00997 Lin. Date: 11/30/2020 Lin. No.: 00031
Solution File No.: 01009 Soln. Date: 03/22/2021 Soln. No.: 00235
Sequential File No.: 01009 File Date: 03/22/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0038
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 0216

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:17S		
Control 1 EC	0.095%	17:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.095%	17:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18S		
Control 2 EC	0.096%	17:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:20S		
Control 3 EC	0.097%	17:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:22S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DOWLING

First Name: NICHOLAS

MI: C

Signature: 

Badge No.: 136

Date: 03/22/2021