Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWM-0030
Location:	DEAL POLICE				
Calibration File No.:	00995	Calib. Date:	11/30/2020	Calib. No.:	00033
Certification File No.:	00996	Cert. Date:	11/30/2020	Cert. No.:	00030
Linearity File No.:	00997	Lin. Date:	11/30/2020	Lin. No.:	00031
Solution File No.:	01011	Soln. Date:	04/06/2021	Soln. No.:	00236
Sequential File No.:	01011	File Date:	04/06/2021		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXA S3-0038
Control Solution %:	0.100%			Expires:	01/13/2023
Solution Control Lot:	21010			Bottle No.:	1003

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:12D		
Control 1 EC	0.097%	01:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	01:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13D		
Control 2 EC	0.099%	01:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:15D		
Control 3 EC	0.099%	01:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO First Name: JONATHAN MI: L

Signature: Pfc. #130 Badge No.: 130

Date: 04/06/2021