

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	00995	Calib. Date: 11/30/2020
Certification File No.:	00996	Calib. No.: 00033
Linearity File No.:	00997	Cert. Date: 11/30/2020
Solution File No.:	01015	Cert. No.: 00030
Sequential File No.:	01015	Lin. Date: 11/30/2020
		Lin. No.: 00031
		Soln. Date: 05/05/2021
		Soln. No.: 00237
		File Date: 05/05/2021

Calibrating Unit:	4	Model No.: CU-34	Serial No.: DDXA S3-0038
Control Solution %:	0.100%		Expires: 01/13/2023
Solution Control Lot:	21010		Bottle No.: 0839

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:30D		
Control 1 EC	-.---%	22:31D	20.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	22:31D	20.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Brad Kerney 131

Badge No.: 131

Date: 05/05/2021