

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWM-0030  
Location: DEAL POLICE  
Calibration File No.: 00995 Calib. Date: 11/30/2020 Calib. No.: 00033  
Certification File No.: 00996 Cert. Date: 11/30/2020 Cert. No.: 00030  
Linearity File No.: 00997 Lin. Date: 11/30/2020 Lin. No.: 00031  
Solution File No.: 01016 Soln. Date: 05/09/2021 Soln. No.: 00238  
Sequential File No.: 01016 File Date: 05/09/2021

Calibrating Unit: 4 Model No.: CU-34 Serial No.: DDXA S3-0038  
Control Solution %: 0.100% Expires: 01/13/2023  
Solution Control Lot: 21010 Bottle No.: 0851

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:39D		
Control 1 EC	0.099%	18:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	18:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:41D		
Control 2 EC	0.099%	18:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	18:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:43D		
Control 3 EC	0.099%	18:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:44D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature:  #130

Badge No.: 130

Date: 05/09/2021