

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWM-0030 |
| Location: | DEAL POLICE | |
| Calibration File No.: | 00995 | Calib. Date: 11/30/2020 |
| Certification File No.: | 00996 | Calib. No.: 00033 |
| Linearity File No.: | 00997 | Cert. Date: 11/30/2020 |
| Solution File No.: | 01017 | Cert. No.: 00030 |
| Sequential File No.: | 01017 | Lin. Date: 11/30/2020 |
| | | Lin. No.: 00031 |
| | | Soln. Date: 05/11/2021 |
| | | Soln. No.: 00239 |
| | | File Date: 05/11/2021 |
| | | |
| Calibrating Unit: | 4 | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXA S3-0038 |
| Solution Control Lot: | 21010 | Expires: 01/13/2023 |
| | | Bottle No.: 0839 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 21:38D | | |
| Control 1 EC | 0.098% | 21:38D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 21:38D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:39D | | |
| Control 2 EC | 0.097% | 21:40D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 21:40D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:41D | | |
| Control 3 EC | 0.098% | 21:42D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 21:42D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:42D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Brad Kerney 131

Badge No.: 131

Date: 05/11/2021