

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	01020	Calib. Date:	05/25/2021
Certification File No.:	01021	Cert. No.:	00034
Linearity File No.:	01022	Cert. Date:	05/25/2021
Solution File No.:	01028	Lin. Date:	05/25/2021
Sequential File No.:	01028	Lin. No.:	00032
		Soln. Date:	06/26/2021
		Soln. No.:	00241
		File Date:	06/26/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0039
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	1006

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:31D		
Control 1 EC	0.099%	22:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:32D		
Control 2 EC	0.099%	22:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34D		
Control 3 EC	0.099%	22:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Brad Kerney B1 DPD

Badge No.: 131

Date: 06/26/2021