

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	01020	Calib. Date: 05/25/2021
Certification File No.:	01021	Calib. No.: 00034
Linearity File No.:	01022	Cert. Date: 05/25/2021
Solution File No.:	01041	Cert. No.: 00031
Sequential File No.:	01041	Lin. Date: 05/25/2021
		Lin. No.: 00032
		Soln. Date: 10/16/2021
		Soln. No.: 00244
		File Date: 10/16/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0039
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 1005

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:48D		
Control 1 EC	0.099%	01:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:50D		
Control 2 EC	0.099%	01:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:51D		
Control 3 EC	0.100%	01:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KERNEY

First Name: BRAD

MI: S

Signature: Sgt. Brad Kerney 131 DPD

Badge No.: 131

Date: 10/16/2021