

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	01020	Calib. Date: 05/25/2021
Certification File No.:	01021	Calib. No.: 00034
Linearity File No.:	01022	Cert. Date: 05/25/2021
Solution File No.:	01046	Cert. No.: 00031
Sequential File No.:	01046	Lin. Date: 05/25/2021
		Lin. No.: 00032
		Soln. Date: 10/24/2021
		Soln. No.: 00246
		File Date: 10/24/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0039
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0990

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:37D		
Control 1 EC	0.099%	22:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38D		
Control 2 EC	0.100%	22:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:40D		
Control 3 EC	0.100%	22:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: DOWLING

First Name: NICHOLAS

MI: C

Signature: _____



Badge No.: 136

Date: 10/24/2021