

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	01048	Calib. Date:	11/12/2021
Certification File No.:	01049	Cert. No.:	00035
Linearity File No.:	01050	Lin. Date:	11/12/2021
Solution File No.:	01054	Cert. No.:	00032
Sequential File No.:	01054	Lin. No.:	00033
		Soln. Date:	11/30/2021
		Soln. No.:	00248
		File Date:	11/30/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0038
Solution Control Lot:	21270	Expires:	07/21/2023
		Bottle No.:	0904

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:56S		
Control 1 EC	0.097%	16:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:58S		
Control 2 EC	0.097%	16:59S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:00S		
Control 3 EC	0.098%	17:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: \_\_\_\_\_

Badge No.: 130

Date: 11/30/2021