

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0030
Location:	DEAL POLICE	
Calibration File No.:	01048	Calib. Date: 11/12/2021
Certification File No.:	01049	Calib. No.: 00035
Linearity File No.:	01050	Cert. Date: 11/12/2021
Solution File No.:	01056	Cert. No.: 00032
Sequential File No.:	01056	Lin. Date: 11/12/2021
		Lin. No.: 00033
		Soln. Date: 12/04/2021
		Soln. No.: 00249
		File Date: 12/04/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0038
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0903

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:02S		
Control 1 EC	0.096%	12:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03S		
Control 2 EC	0.097%	12:04S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		
Control 3 EC	0.097%	12:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: \_\_\_\_\_

Badge No.: 130

Date: 12/04/2021