

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWM-0030
Location:	DEAL POLICE		
Calibration File No.:	01020	Calib. Date:	05/25/2021
Certification File No.:	01021	Cert. No.:	00034
Linearity File No.:	01022	Cert. Date:	05/25/2021
Solution File No.:	01036	Lin. No.:	00032
Sequential File No.:	01036	Lin. Date:	05/25/2021
		Soln. Date:	08/23/2021
		Soln. No.:	00242
		File Date:	08/23/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXA S3-0039
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	0993

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:24D		
Control 1 EC	0.099%	20:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:26D		
Control 2 EC	0.099%	20:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27D		
Control 3 EC	0.100%	20:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DOWLING

First Name: NICHOLAS

MI: C

Signature: \_\_\_\_\_

*[Handwritten Signature]* #136

Badge No.: 136

Date: 08/23/2021