Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARWM-0030
Location:	DEAL POLICE				
Calibration File No.:	01048	Calib. Date:	11/12/2021	Calib. No.:	00035
Certification File No.:	01049	Cert. Date:	11/12/2021	Cert. No.:	00032
Linearity File No.:	01050	Lin. Date:	11/12/2021	Lin. No.:	00033
Solution File No.:	01061	Soln. Date:	01/17/2022	Soln. No.:	00251
Sequential File No.:	01061	File Date:	01/17/2022		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXA S3-0038
Control Solution %:	0.100%			Expires:	07/21/2023
Solution Control Lot:	21270			Bottle No.:	1124

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:45S		
Control 1 EC	0.101%	01:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:47S		
Control 2 EC	0.101%	01:48S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:49S		
Control 3 EC	0.101%	01:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO First Name: JONATHAN MI: L

Signature: #130 Badge No.: 130

Date: 01/17/2022