

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWM-0030 |
| Location: | DEAL POLICE | |
| Calibration File No.: | 01066 | Calib. Date: 03/15/2022 |
| Certification File No.: | 01067 | Calib. No.: 00036 |
| Linearity File No.: | 01068 | Cert. Date: 03/15/2022 |
| Solution File No.: | 01071 | Cert. No.: 00033 |
| Sequential File No.: | 01071 | Lin. Date: 03/15/2022 |
| | | Lin. No.: 00034 |
| | | Soln. Date: 04/25/2022 |
| | | Soln. No.: 00254 |
| | | File Date: 04/25/2022 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXA S3-0039 |
| Solution Control Lot: | 21270 | Expires: 07/21/2023 |
| | | Bottle No.: 1039 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 00:37D | | |
| Control 1 EC | 0.101% | 00:38D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 00:38D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 00:38D | | |
| Control 2 EC | 0.101% | 00:39D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 00:39D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 00:40D | | |
| Control 3 EC | 0.101% | 00:41D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 00:41D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 00:42D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature: _____

Badge No.: 130

Date: 04/25/2022