

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0030
Location: DEAL POLICE
Calibration File No.: 01066 Calib. Date: 03/15/2022 Calib. No.: 00036
Certification File No.: 01067 Cert. Date: 03/15/2022 Cert. No.: 00033
Linearity File No.: 01068 Lin. Date: 03/15/2022 Lin. No.: 00034
Solution File No.: 01074 Soln. Date: 05/31/2022 Soln. No.: 00255
Sequential File No.: 01074 File Date: 05/31/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0039
Control Solution %: 0.100% Expires: 07/21/2023
Solution Control Lot: 21270 Bottle No.: 0777

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:29D		
Control 1 EC	0.102%	13:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31D		
Control 2 EC	0.101%	13:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33D		
Control 3 EC	0.102%	13:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LOSQUADRO

First Name: JONATHAN

MI: L

Signature:  #130

Badge No.: 130

Date: 05/31/2022